


# Declaration of Conformity

<b>Applications of Council Directive: 93/42/EEC Medical Devices Directive</b>	
Standards to which conformity is declared	EN 60601-1, IEC 60601-2-30, EN 60601-1-2 (EMC), EN 1060-1, EN 1060-3, “Non-Invasive Sphygmomanometers – General Requirements & Supplementary Requirements for Electro-Mechanical BP Measuring Systems”, AAMI SP10
Manufacturer’s name and address	SunTech Medical, Inc. 507 Airport Boulevard, Suite 117 Morrisville, NC 27560-8200 USA
Importer’s name and address	SunTech Medical, Ltd. Europe Oakfield Industrial Estate Stanton Harcourt Rd. Eynsham, Oxfordshire OX29 4TS England
Type of Equipment	Accutracker® Ambulatory Blood Pressure Monitor and report generator software Class IIa
Model Number	Accutracker II™, Accutracker DX™

I, the undersigned, hereby declare that the equipment specified above conforms to the above Directive(s) and Standard(s).



DECEMBER 9, 2004

Name: Dayn McBee  
Position: Chief Executive Officer  
Company: SunTech Medical, Inc.

Date